



Certificate of Compliance

Certificate: 80033020

Master Contract: 213862

Project: 80033020

Date Issued: 2020-02-18

Issued To: Vaisala Oyj
PO Box 26
Helsinki, ., 00421
Finland

Attention: J-P Mörsky

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: *Sam Zaffino*
Sam Zaffino



PRODUCTS

CLASS - C225205 - PROCESS CONTROL EQUIPMENT

CLASS - C225285 - PROCESS CONTROL EQUIPMENT Certified to US Standards

PR-23 series inline refractometer (4-20mA); and DTR (indicating transmitter), permanently connected, rated 100...240V, 50/60Hz, 30VA.

Note: The above model is Equipment Class I, Pollution Degree 2 (DTR) / Pollution Degree 3 (PR-23 series), Installation Category II.



Certificate: 80033020
Project: 80033020

Master Contract: 213862
Date Issued: 2020-02-18

APPLICABLE REQUIREMENTS

- CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements
- UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 80033020

Master Contract: 213862

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80033020	2020-02-18	Transfer of 1688113 and to cover alternate construction